

EX08-005

**HAWAII HEALTH SYSTEMS CORPORATION
NOTICE OF AND REQUEST FOR EXEMPTION
FROM CHAPTER 103D, HRS**

- 1. TO: Chief Procurement Officer
- 2. FROM: Arthur Sampaga, Nursing Department

Department/Division/Agency

Pursuant to §103D-102(b)(4), HRS, and Chapter 3-120, HAR, the Department requests a procurement exemption to purchase the following:

3. Description of goods, services or construction:

Axiom Sensis Hemodynamic monitoring system is a state of the art equipment needed to perform Cardiac Catheterization procedures. The Sensis system records all types of hemodynamic studies with a built in Syngo database for long term storage of patient's examinations including measurements, waveforms, event logs, and flowsheet. The complete software is for all-pediatric / adult, right/left heart, angio/valvular hemodynamic examinations including functionality for calculations such as gradients, valve areas, shunts. Including annotation, event log, split screen and 12 channel ECG's. This system is primarily configured to the present Siemens Multistar Angiography machine. Upgrading would take approximately 3 days. Basic configuration of the Sensis system will include; 2 FSC PC in a cabinet with 1.7 GHz Pentium 4 processor, 1024 MB RAM, 40 GB Hard disk, CD R-writer, and Tape drive.

4. Name of Vendor: Siemens Medical Solutions USA, Inc. Address: 670 Koll Center Parkway, Suite 220 Pleasanton, CA 94566	5. Price: \$171,225.00
---	---------------------------

6. Term of Contract: From: Execution date To: One year from date of execution	7. Prior Exemption Ref. No.
---	-----------------------------

8. Explanation describing how procurement by competitive means is either not practicable or advantageous to the State:

be The Axiom Sensis Hemodynamic system we specifically design for this type of upgrade, which compliments HMC's present Siemens Multistar Angiography machine. Sensis and Siemens have been integrated systems for a while now. A major advantages will that the two systems function well together, one service person from Siemens is all that is needed instead of dealing with two separate vendors, thus minimizing any disruption in patient services. Even beyond the Safety, Service, and Cost considerations noted for the Axiom Sensis Physiological monitoring system, the technology platform and custom design would not be practical for completion based upon the existing Siemens Axiom Artis Imaging System interface required.

9. Details of the process or procedures to be followed in selecting the vendor to ensure maximum fair and open competition as practicable:

Axiom Sensis Hemodynamic system was made specifically for our present Siemens Angiography machine. The Sensis vendor ensures fully streamline configuration between two systems that will communicate between each other without difficulty.

Siemens has a service technician on Island that could respond to immediate trouble shoot calls for both Siemens and Sensis systems. In his absence, a temporary person will be relocated on Island. Other vendors do not offer immediate service calls. Their service personnel are on Oahu and will take hours if not days to respond to HMC.

The cost is kept at a minimum since both systems are made to configure with each other.

10. A description of the agency's internal controls and approval requirements for the exempted procurement:
 Director of Contracts and Material Mgt. Reviews request for exemption for sufficiency prior to submission to the President and CEO for action. HHSC Policies and Procedures will be followed and periodic reviews of contract action will be conducted.


11. A list of agency personnel, by position, who will be involved in the approval process and administration of the contract:

Name	Position	Involvement in Process	
Ron Schurra	Regional CEO	<input checked="" type="checkbox"/> Approval	<input type="checkbox"/> Administration
Money Atwal	Regional CFO	<input type="checkbox"/> Approval	<input checked="" type="checkbox"/> Administration
Julie-Beth Ako	Director of Clinical Support	<input type="checkbox"/> Approval	<input checked="" type="checkbox"/> Administration
Dan Brinkman	Chief Nurse Executive	<input type="checkbox"/> Approval	<input checked="" type="checkbox"/> Administration
Arthur Sampaga	Cardiovascular Supervisor	<input type="checkbox"/> Approval	<input checked="" type="checkbox"/> Administration
Debbie Kaina	Regional Materials Manager	<input type="checkbox"/> Approval	<input checked="" type="checkbox"/> Administration

12. Direct inquiries to:
 Department: Angiography
 Contact Name: Arthur Sampaga, RN
 Phone Number: 808-974-6950
 Fax Number: 808-933-0509

13. This exemption should be considered for list of exemptions attached to Chapter 3-120, HAR: YES NO
 Agency shall ensure adherence to applicable administrative and statutory requirements

14. I certify that the information provided above is, to the best of my knowledge, true and correct.


 Department Head

13 Sep 07
 Date


Reserved for CFO/Delegee Use Only

15. Date Notice Posted 10-8-07

The Head of the Purchasing Agency is in the process of reviewing this request for exemption from Chapter 103D, HRS. Submit written objections to this notice to issue an exemption from Chapter 103D, HRS, within seven (7) calendar days or as otherwise allowed from the above posted date to:
 Chief Procurement Officer
 HHSC
 3675 Kilauea Ave.
 Honolulu, Hawaii 96816

Chief Procurement Officer's comments:

16. APPROVED DISAPPROVED


 Chief Procurement Officer
 10/19/07
 Date